

# GAS INNOVATIONS

## CERTIFICATE OF CONFORMANCE

### ETHANE

### C. P. GRADE

<u>Guaranteed Specifications</u>		<u>Minimal Purity</u>
99.0 % +	Ethane	99.0 % +
Sum of all impurities to be < 1 %	Propane	< 0.60 %
	Methane	< 0.39 %
Sulfur < 1 PPM	Nitrogen	< 100 PPM
Water < 3 PPM	Oxygen	< 20 PPM
	Other Hydrocarbons	< 10 PPM
	CO + CO2	< 20 PPM
	Water	< 2 PPM
	Sulfur	< 0.50 PPM

Certified by: \_\_\_\_\_

Date: \_\_\_\_\_

\*\* Liquid Phase Analysis – Reported In Mole Per Cent \*\*

SO# 00

Cylinder Serial Number(s)  
**Size K:**

Batch Number  
ETA-051018-C

CUSTOMER PO#: **000**

Lot Number: ETA-051018-TT-085175-4

18005 E. Highway 225, La Porte, TX 77571 Tel- 281-471-2200 Fax- 281-471-2201 [www.gasinnovations.com](http://www.gasinnovations.com)