

GAS INNOVATIONS

CERTIFICATE OF CONFORMANCE

ISOBUTANE 3.0 UHP PURITY

<u>Guaranteed Specifications</u>		<u>Minimal Purity</u>
99.9 % +	Isobutane	99.90 % +
Sum of all impurities to be < 0.10 %	Methane	< 1 PPM
	Ethane	< 1 PPM
	Propane	< 300 PPM
Water < 2 PPM	n-Butane	< 600 PPM
Liquid Phase Analysis	Isopentane	< 100 PPM
	Water	< 1 PPM
	Sulfur	< 0.1 PPM

Certified by: _____

Date: _____

SO# 00

Cylinder Serial Number(s)

Batch Number

IBU-112817-UHP

CUSTOMER PO#: **000**

Lot Number: IBU-112817-UHP

18005 E. Highway 225, La Porte, TX 77571 Tel-281-471-2200 Fax-281-471-2201 www.gasinnovations.com