

GAS INNOVATIONS

CERTIFICATE OF CONFORMANCE

CARBON MONOXIDE C.P. GRADE

Guaranteed Specifications

Minimal Purity

99.5 % +	Carbon Monoxide	99.95 % +
Sum of all impurities to be < 0.50 %	Argon	< 200 PPM
Water < 10 PPM	Hydrogen	< 10 PPM
	Nitrogen	< 2500 PPM
	Oxygen	< 5 PPM
	THC (as Methane)	< 15 PPM
	Carbon Dioxide	< 25 PPM
	Water	< 5 PPM

Certified by: _____

Date: _____

SO# 00

Cylinder Serial Number(s)

Batch Number

Size T:

CO-120418-C

CUSTOMER PO#: **000**

Lot Number: CO-111218-TT-1-10-1056

18005 E. Highway 225, La Porte, TX 77571 Tel- 281-471-2200 Fax- 281-471-2201 www.gasinnovations.com