

GAS INNOVATIONS

CERTIFICATE OF CONFORMANCE

CARBON MONOXIDE RESEARCH GRADE

<u>Guaranteed Specifications</u>		<u>Minimal Purity</u>
99.99 % +	Carbon Monoxide	99.99% +
Sum of all impurities to be < 0.01 %	Argon	< 10 PPM
	Hydrogen	< 25 PPM
Water < 5 PPM	Nitrogen	< 60 PPM
	Oxygen	< 0.1 PPM
	THC (as Methane)	< 1 PPM
	Carbon Dioxide	< 5 PPM
	Water	< 1 PPM

Certified by: _____

Date: _____

SO# 00

Cylinder Serial Number(s)
Size **150AL**:

Batch Number
CO-071018-R

CUSTOMER PO#:

Lot Number: CO-071018-eCOs

18005 E. Highway 225, La Porte, TX 77571 Tel- 281-471-2200 Fax- 281-471-2201 www.gasinnovations.com