

GAS INNOVATIONS

CERTIFICATE OF CONFORMANCE

ETHANE 4.0 UHP+ GRADE

<u>Guaranteed Specifications</u>		<u>Minimal Purity**</u>
99.99 % +	Ethane	99.995 % +
Sum of all impurities to be < 100 ppmv	Ethylene	< 0.1 ppmv
	Propane	< 25 ppmv
	Butanes	< 0.1 ppmv
	Pentanes	< 0.1 ppmv
	CO + CO2	< 0.1 ppmv
	Hydrogen	< 15 ppmv
	Methane	< 1 ppmv
	Nitrogen	< 5 ppmv
	Oxygen	< 1 ppmv
	Water	1 ppmv

Certified by: _____

Date: _____

**** Vapor Phase Analysis ****

SO# 00

Cylinder Serial Number(s)

Batch Number
ETA-072618-UHP+

CUSTOMER PO#: **000**

Lot Number: ETA-072318-TT-085196-5

18005 E. Highway 225, La Porte, TX 77571 Tel- 281-471-2200 Fax- 281-471-2201 www.gasinnovations.com