

GAS INNOVATIONS

CERTIFICATE OF CONFORMANCE

ETHANE

INSTRUMENT GRADE

Guaranteed Specifications

Lot Analysis

99.5 % +	Ethane	99.63 % +
Sum of all impurities to be < 0.5 %	Propane	0.15 %
Sulfur < 1 PPM	Methane	0.23 %
Water < 3 PPM	Butanes	<5 PPM
	Pentanes	<1 PPM
	Carbon Dioxide	< 100 PPM
	Water	< 2 PPM
	Sulfur	< 0.50 PPM

Certified by: _____

Date: _____

**** Liquid Phase Analysis – Reported In Mole Per Cent ****

Cylinder Serial Number(s)

Batch Number
ETA-093013-I

Lot Number: ETA-092313TTWS220-7