

GAS INNOVATIONS

CERTIFICATE OF CONFORMANCE

ETHYLENE 3.5 U.H.P. PLUS GRADE

Guaranteed Specifications

99.95 % +

Sum of all impurities to

be < 0.05 %

Ethylene

Acetylene

Carbon Dioxide

Carbon Monoxide

Ethane

Methane

Nitrogen

Water

Sulfur

Minimal Purity

99.96 % +

< 1 PPM

< 1 PPM

< 1 PPM

< 350 PPM

< 8 PPM

< 10 PPM

< 1 PPM

< 0.01 PPM

Certified by: _____

Date: _____

SO# 00

Cylinder Serial Number(s)

Size T:

Batch Number

ETY-032118-UHP+

CUSTOMER PO#: **000**

Lot Number: ETY-032118-STK-11K