

GAS INNOVATIONS

CERTIFICATE OF CONFORMANCE

ISOBUTYLENE

3.0

UHP PURITY

<u>Guaranteed Specifications</u>		<u>Minimal Purity</u>
99.9 % +	Isobutylene	99.90 % +
Sum of all impurities to be < 0.10 %	Butene-1	< 10 PPM
	2-Butenes	< 10 PPM
Sulfur < 0.50 PPM	1,3 Butadiene	< 40 PPM
Water < 3 PPM	n-Butane	< 5 PPM
	Isobutane	< 50 PPM
Liquid Phase Analysis	Propane	< 5 PPM
	Propylene	< 1 PPM
	Water	< 3 PPM
	Sulfur	< 0.50 PPM

Certified by: _____

Date: _____

SO# 000

Cylinder Serial Number(s)

Batch Number
IBY-110717-UHP

Customer PO# :

Lot Number: IBY-110717-UHP

18005 E. Highway 225, La Porte, TX 77571 Tel- 281-471-2200 Fax- 281-471-2201 www.gasinnovations.com