

# GAS INNOVATIONS

## CERTIFICATE OF CONFORMANCE

### ISOBUTYLENE

### INSTRUMENT PURITY

#### Guaranteed Specifications

99.5 % +

Sum of all impurities to

be < 0.50 %

Sulfur < 0.50 PPM

Water < 3 PPM

#### Liquid Phase Analysis

#### Minimal Purity

99.5 % +

< 1 PPM

< 1 PPM

< 1 PPM

< 100 PPM

< 200 PPM

< 5 PPM

< 1 PPM

< 2 PPM

< 0.50 PPM

Certified by: \_\_\_\_\_

Date: \_\_\_\_\_

SO# 000

Cylinder Serial Number(s)

Batch Number

IBY-011217-I

Customer PO#: **000**

Lot Number: IBY-011217-I