

GAS INNOVATIONS

CERTIFICATE OF CONFORMANCE

METHANE, CP. GRADE

<u>Guaranteed Specifications</u>		<u>Minimal Purity</u>
99.5 % +	Methane	99.5 % +
Sum of all impurities to be < 0.5 %	Ethane	< 4000 PPM
	Oxygen /Argon	< 100 PPM
	Nitrogen	< 1,000 PPM
	Carbon Dioxide	< 30 PPM
	Other Hydrocarbons (Total)	< 1,000 PPM
	Carbon Monoxide	< 5 PPM
	Water	< 5 PPM

Certified by: _____

Date: _____

Cylinder Serial Number(s)

Batch Number
MET-012113-C

Lot Number: MET-012113-STK

18005 E. Highway 225, La Porte, TX 77571 Tel- 281-471-2200 Fax- 281-471-2201 www.gasinnovations.com